


Test Bank for Power Instrumentation for the Dental Professional 1st Edition by Mayo

[CLICK HERE TO ACCESS COMPLETE Test Bank](#)



Test Bank

Power Instrumentation for the Dental Professional, First Edition

Lisa Mayo, RDH, BSDH, MHA

Test Bank

Chapter 2

Multiple Choice

1. Which of the following is a reason why a biofilm can readily bind to and build colonies inside dental unit water line tubing?

- A) Tubing is plastic with a narrow lumen size
- B) Tubing has inconsistent water flow rates
- C) Water can become stagnant inside the tubing
- D) All of the above

Answer: D

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Moderate

Bloom's Taxonomy: Remembering/Understanding

2. Which of the following devices will remove 100 cubic of air per minute and reduce environmental aerosol contamination by at minimum 90% when used as instructed by the manufacturer?

- A). Low volume evacuation
- B.) Extraoral scavenging device
- C.) Filtering facepiece respirator
- D.) High volume evacuation

Answer: D

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Moderate

Bloom's Taxonomy: Applying/Analyzing

3. Which of the following is a technology that can improve indoor air quality and reduce the risk of disease transmission from contaminated air?

- A). UV light
- B.) Vapor hydrogen peroxide
- C.) High volume evacuation
- D.) Both A & B
- E.) All of the above

Answer: D

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Moderate

Bloom's Taxonomy: Applying/Analyzing

True/False

Power Instrumentation for the Dental Professional, First Edition

Lisa Mayo, RDH, BSDH, MHA

Test Bank

1. True or False? A biofilm is a collection of microbial cells that bind together on a surface and form an enclosed matrix held together by polysaccharides that is reversibly bound to a surface.

Answer: False

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Easy

Bloom's Taxonomy: Remembering/Understanding

2. True or False? Water and vacuum tubing on a dental unit does not provide a favorable environment for the growth of a biofilm.

Answer: False

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Easy

Bloom's Taxonomy: Remembering/Understanding

3. True or False? According to the CDC, it is acceptable for the dental unit water line to have 250CFU/mL heterotrophic mesophilic water bacteria in the line for a surgical procedure

Answer: False

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Easy

Bloom's Taxonomy: Remembering/Understanding

4. True or False? According to the CDC, it is acceptable for the dental unit water line to have 400CFU/mL heterotrophic mesophilic water bacteria in the line for a non-surgical procedure.

Answer: True

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Easy

Bloom's Taxonomy: Remembering/Understanding

5. True or False? The ASTM has set three performance levels for filtering facepiece respirator masks.

Answer: False

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Easy

Bloom's Taxonomy: Remembering/Understanding

6. True or False? If you wear a face shield during an aerosol-generating procedure, you do not need to additionally wear protective eye wear.

Answer: False

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Easy

Power Instrumentation for the Dental Professional, First Edition

Lisa Mayo, RDH, BSDH, MHA

Test Bank

Bloom's Taxonomy: Remembering/Understanding

7. True or False? A surgical mask will provide higher protection for the wearer from small particle size aerosols compared to a filtering facepiece respirator mask.

Answer: False

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Easy

Bloom's Taxonomy: Remembering/Understanding

8. True or False? If you have a HEPA filter in your dental operatory, you do not have to use a high volume evacuation device during a dental aerosol-generating procedure.

Answer: False

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Moderate

Bloom's Taxonomy: Evaluating/Creating

Multiple Response

1. Which of the following would you prefer to use during a dental aerosol-generating procedure that offers the highest level of protection for the oral health care provider? Select all that apply.

A. Surgical mask

B. Filtering facepiece respirator mask

C. Low volume evacuation

D. High volume evacuation

E. Hair covering

F. Face shield

Answer: B, D, E, F

A-Head: Dental aerosol management

Subject: Chapter 2

Complexity: Moderate

Bloom's Taxonomy: Applying/Analyzing

2. Which of the following is an agent present in alcohol-based hand antiseptics used in health care? Select all that apply.

A. Chlorhexidine gluconate

B. Hexachlorophene

C. Triclosan

D. Isopropanol

E. Ethanol

F. Chloroxylonol

Answer: A, B, C, D, E

A-Head: Dental aerosol management

Subject: Chapter 2

Power Instrumentation for the Dental Professional, First Edition

Lisa Mayo, RDH, BSDH, MHA

Test Bank

Complexity: Difficult

Bloom's Taxonomy: Remembering/Understanding